

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Lemo21(DE3) Competent E. coli

Catalog Number: C2528J
Lot Number: 10048002
Expiration Date: 05/2020
Storage Temperature: -80°C

Specification Version: PS-C2528H v1.0

Lemo21(DE3) Competent E. coli Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N3041AVIAL	pUC19 Vector	10043515	Pass	
C2528JVIAL	Lemo21(DE3) Competent E. coli	10029095	Pass	
B9030AVIAL	L-rhamnose solution	10014840	Pass	

Assay Name/Specification	Lot # 10048002
Antibiotic Sensitivity (Tetracycline) 15 µl of untransformed Lemo21(DE3) Competent E. coli streaked onto a Rich Broth plate containing Tetracycline will not form colonies after incubation for 16 hours at 37°C.	Pass
Antibiotic Resistance (Chloramphenicol) 15 µl of untransformed Lemo21(DE3) Competent E. coli streaked onto a Rich Broth plate containing Chloramphenicol will form colonies after incubation for 16 hours at 37°C.	Pass
Antibiotic Sensitivity (Ampicillin) 15 µl of untransformed Lemo21(DE3) Competent E. coli streaked onto a Rich Broth plate containing Ampicillin will not form colonies after incubation for 16 hours at 37°C.	Pass
Antibiotic Sensitivity (Kanamycin) 15 µl of untransformed Lemo21(DE3) Competent E. coli streaked onto a Rich Broth plate containing Kanamycin will not form colonies after incubation for 16 hours at 37°C.	Pass
Antibiotic Sensitivity (Spectinomycin) 15 µl of untransformed Lemo21(DE3) Competent E. coli streaked onto a Rich Broth plate containing Spectinomycin will not form colonies after incubation for 16 hours	Pass



C2528J / Lot: 10048002

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Lixin An

Production Scientist

20 Nov 2018

Corey Rabeau

Packaging Quality Control Inspector

13 Jun 2019



C2528J / Lot: 10048002

Page 2 of 2