

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEB® PCR Cloning Kit

Catalog Number: E1202S
Packaging Lot Number: 10082505
Expiration Date: 05/2021

Storage Temperature: Multi-temperature\*
Specification Version: PS-E1202S v2.0

\* This product contains components with different storage temperature requirements. Please reference the applicable product specification document(s) on the Quality and Safety tab located on the product page of www.neb.com.

NEB® PCR Cloning Kit Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
S1513AVIAL	Cloning Analysis Reverse Primer	10071361	Pass	
S1512AVIAL	Cloning Analysis Forward Primer	10071360	Pass	
N3041AVIAL	pUC19 Vector	10078142	Pass	
N0555AVIAL	Amplicon Cloning Control	10071359	Pass	
N0312AVIAL	Linearized pMiniT™ 2.0 Vector	10071358	Pass	
M1016AVIAL	Cloning Master Mix 2	10071357	Pass	
M1015AVIAL	Cloning Master Mix 1	10071356	Pass	
C3018AVIAL	NEB® 10-beta Competent E. coli (Cloning Efficiency)	10078639	Pass	
B9035SVIAL	NEB® 10-beta/Stable Outgrowth Medium	10079002	Pass	

Assay Name/Specification	Lot # 10082505
Functional Testing (NEB® PCR Cloning Kit)	Pass
A 10 µl ligation reaction containing 5 µl cloning mix, 25 ng of linearized pMiniT™	
2.0 vector and 30 ng of a 1 kb amplicon containing A overhangs is incubated for 5	
minutes at 25°C. Transformation of NEB 10-beta Competent E. coli with 2 µl of the	
ligated master mix/fragment mixture yields ≥ 100 colonies when 5% of the	
transformation is plated on an ampicillin plate and incubated overnight at 37°C. The	
colonies are screened by colony PCR and ≥95% of colonies have an insert present.	
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in NEB® PCR	
Cloning Kit and meet the designated specifications.	



E1202S / Lot: 10082505

Page 1 of 2



This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Ana Egana
Production Scientist

17 Aug 2020

Corey Rabeau

Packaging Quality Control Inspector

17 Aug 2020



E1202S / Lot: 10082505

Page 2 of 2