

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEB® PCR Cloning Kit (Without Competent Cells)

Catalog Number: E1203S
Packaging Lot Number: 10176797
Expiration Date: 01/2024
Storage Temperature: -20°C

Specification Version: PS-E1203S v1.0

NEB® PCR Cloning Kit (Without Competent Cells) Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
S1513AVIAL	Cloning Analysis Reverse Primer	10169109	Pass	
S1512AVIAL	Cloning Analysis Forward Primer	10169108	Pass	
N0555AVIAL	Amplicon Cloning Control	10169107	Pass	
N0312AVIAL	Linearized pMiniT™ 2.0 Vector	10169106	Pass	
M1016AVIAL	Cloning Master Mix 2	10169105	Pass	
M1015AVIAL	Cloning Master Mix 1	10169104	Pass	

Assay Name/Specification	Lot # 10176797
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEB® PCR	Pass
Cloning Kit (Without Competent Cells) and meet the designated specifications.	
Functional Testing (NEB® PCR Cloning Kit)	Pass
A 10 µl ligation reaction containing 5 µl cloning mix, 25 ng of linearized pMiniT™	
2.0 vector and 30 ng of a 1 kb amplicon containing A overhangs is incubated for 5	
minutes at 25°C. Transformation of NEB 10-beta Competent E. coli with 2 μl of the	
ligated master mix/fragment mixture yields ≥ 100 colonies when 5% of the	
transformation is plated on an ampicillin plate and incubated overnight at 37°C. The	
colonies are screened by colony PCR and ≥95% of colonies have an insert present.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E1203S / Lot: 10176797 Page 1 of 2 Lindsey Spiegelman Production Scientist 24 Jan 2023

Michael Tonello

Packaging Quality Control Inspector

06 Feb 2023