

New England Biolabs Certificate of Analysis

Product Name: NEBNext[®] Ultra[™] II Ligation Module
Catalog Number: E7595S
Packaging Lot Number: 10109302
Expiration Date: 05/2022
Storage Temperature: -20°C
Specification Version: PS-E7595S/L v1.0

NEBNext [®] Ultra [™] II Ligation Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7648AVIAL	NEBNext [®] Ultra [™] II Ligation Master Mix	10109304	Pass
E7374AVIAL	NEBNext [®] Ligation Enhancer	10109303	Pass

Assay Name/Specification	Lot # 10109302
Functional Testing (Library Construction, DNA Modules) Each set of reagents is functionally validated and compared to a previous lot through construction of DNA libraries. Libraries made from previous and current lots for the recommended input DNA amounts are sequenced together on the same Illumina [®] flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext [®] Ultra [™] II Ligation Module and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christine Sumner

Christine Sumner
Production Scientist
27 May 2021

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
27 May 2021