

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: LongAmp® Taq 2X Master Mix

Catalog Number: M0287L

Concentration: 2 X Concentrate

Packaging Lot Number: 10071793
Expiration Date: 09/2021
Storage Temperature: -20°C

Specification Version: PS-M0287S/L v2.0

Composition (1X): 60 mM Tris-SO4 (pH 9.1 @ 25°C), 20 mM (NH4)2SO4, 2 mM MgSO4, 0.3 mM

dATP, 0.3 mM dCTP, 0.3 mM dGTP, 0.3 mM dTTP, 3 % Glycerol, 0.06 % IGEPAL® CA-630, 0.05 % Tween® 20, 125 units/ml LongAmp® Taq DNA

Polymerase

| LongAmp® Taq 2X Master Mix Component List | | | | |
|---|----------------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| M0287SVIAL | LongAmp® Taq 2X Master Mix | 10069425 | Pass | |

| Assay Name/Specification | Lot # 10071793 |
|---|----------------|
| Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X LongAmp® Taq Master Mix containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| PCR Amplification (30 kb Human Genomic DNA, Master Mix) A 25 µl reaction in 1X LongAmp® Taq Master Mix and 0.4 µM primers containing 500 ng Human Genomic DNA for 28 cycles of PCR amplification results in the expected 30 kb product. | Pass |
| PCR Amplification (30 kb Lambda DNA, Master Mix) A 25 µl reaction in 1X LongAmp® Taq Master Mix and 0.4 µM primers containing 1 ng Lambda DNA for 28 cycles of PCR amplification results in the expected 30 kb product. | Pass |
| qPCR DNA Contamination (E. coli Genomic) A minimum of 2.5 units of LongAmp® Taq DNA Polymerase is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome. | Pass |



M0287L / Lot: 10071793

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Cathy Rezac Production Scientist

31 Mar 2020

Josh Hersey

Packaging Quality Control Inspector

31 Mar 2020

