

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Poly (U) Polymerase

Catalog Number: M0337S
Concentration: 2,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that incorporates 1 nmol

of UMP into RNA in a 50 µl reaction incubated for 10 minutes at

37°C.

Packaging Lot Number: 10065475
Expiration Date: 04/2022
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 100 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol,

(pH 7.5 @ 25°C)

Specification Version: PS-M0337S v1.0

Poly (U) Polymerase Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0337SVIAL	Poly (U) Polymerase	10072477	Pass	
B7002SVIAL	NEBuffer™ 2	10061303	Pass	

Assay Name/Specification	Lot # 10065475
Endonuclease Activity (Nicking) A 50 μl reaction in NEBuffer 2 containing 1 μg of supercoiled pUC19 DNA and a	Pass
minimum of 10 units of Poly(U) Polymerase incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 2 containing 1 µg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 10 units of Poly(U) Polymerase incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Protein Purity Assay (SDS-PAGE) Poly(U) Polymerase is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass
and a minimum of 2 units of Poly(U) Polymerase is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by	



M0337S / Lot: 10065475

Page 1 of 2

Assay Name/Specification	Lot # 10065475
gel electrophoresis using fluorescent detection	

This product has been tested and shown to be in compliance with all specifications.

Bhairavi Jani Production Scientist

03 Apr 2020

Jay Minichiello

Packaging Quality Control Inspector

03 Apr 2020