

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTag® Hot Start 2X Master Mix with Standard Buffer

Catalog Number: M0484S

Concentration: 2 X Concentrate

Packaging Lot Number: 10156553
Expiration Date: 02/2024
Storage Temperature: -20°C

Specification Version: PS-M0484S/L v2.0

Composition (1X): 20 mM Tris-HCl (pH 8.9 @ 25°C), 22 mM NH4Cl, 22 mM KCl, 1.8 mM

MgCl2, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 5 % Glycerol, 0.06 % IGEPAL® CA-630, 0.05 % Tween® 20, 25 units/ml

OneTaq® Hot Start DNA Polymerase

OneTaq® Hot Start 2X Master Mix with Standard Buffer Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0484SVIAL	OneTaq® Hot Start 2X Master Mix with Standard Buffer	10135597	Pass	

Assay Name/Specification	Lot # 10156553
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X OneTaq® Hot Start Master Mix with Standard Buffer containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
PCR Amplification (Hot Start 2 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 pg Lambda DNA and 50 ng Human Genomic DNA with 0.625 units of OneTaq® Hot Start DNA Polymerase for 30 cycles of PCR amplification results in an increase in yield of the 2 kb Lambda product and a decrease in non-specific genomic bands when compared to a non-hot start control reaction.	Pass
PCR Amplification (5 kb Lambda, Master Mix) A 25 μl reaction in 1X OneTaq® Hot Start Master Mix with Standard Buffer and 0.2 μM primers containing 5 ng Lambda DNA for 25 cycles of PCR amplification results in the expected 5 kb product.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass



M0484S / Lot: 10156553

Page 1 of 2

>95% inhibition when compared to a non-hot start control reaction.

vistie Vanguez

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 23 Jun 2022 Erin Varney

Packaging Quality Control Inspector

23 Jun 2022



M0484S / Lot: 10156553

Page 2 of 2