

New England Biolabs Certificate of Analysis


Product Name: Phusion® High-Fidelity PCR Master Mix with GC Buffer
Catalog Number: M0532S
Concentration: 2 X Concentrate
Packaging Lot Number: 10218712
Expiration Date: 03/2025
Storage Temperature: -20°C
Specification Version: PS-M0532S/L v2.0
Composition (1X): 1 X Phusion® GC Buffer, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 20 units/ml Phusion® High-Fidelity DNA Polymerase

Phusion® High-Fidelity PCR Master Mix with GC Buffer Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0532SVIAL	Phusion® High-Fidelity PCR Master Mix with GC Buffer	10209995	Pass
B0515AVIAL	DMSO	10194246	Pass

Assay Name/Specification	Lot # 10218712
PCR Amplification (20 kb Lambda DNA, Master Mix) A 50 µl reaction in 1X Phusion® High-Fidelity PCR Master Mix with GC Buffer and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.	Pass
PCR Amplification (7.5 kb Human Genomic DNA, Master Mix) A 50 µl reaction in 1X Phusion® High-Fidelity PCR Master Mix with GC Buffer and 1.0 µM primers containing 50 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7.5 kb product.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.


Lea Antonopoulos
Production Scientist
25 Oct 2023


Josh Hersey
Packaging Quality Control Inspector
07 Dec 2023