

New England Biolabs Certificate of Analysis


Product Name: MspI
Catalog Number: R0106S
Concentration: 20,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10091169
Expiration Date: 12/2021
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0106S/L v1.0

| MspI Component List | | | |
|---------------------|------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| R0106SVIAL | MspI | 10061361 | Pass |
| B7204SVIAL | CutSmart® Buffer | 10089399 | Pass |
| B7024AVIAL | Gel Loading Dye, Purple (6X) | 10084973 | Pass |


| Assay Name/Specification | Lot # 10091169 |
|---|----------------|
| Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with MspI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with MspI. | Pass |
| Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of MspI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of MspI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity. | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Mala Samaranayake
Production Scientist
18 Nov 2020



Michael Tonello
Packaging Quality Control Inspector
18 Nov 2020