

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Nsil-HF®
Catalog Number: R3127L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10059959
Expiration Date: 11/2021
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl , 10 mM Tris-HCl (pH 7.4), 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol , 500 μg/ml BSA

Specification Version: PS-R3127S/L v1.0

Nsil-HF® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R3127LVIAL	Nsil-HF®	10059957	Pass	
B7204SVIAL	CutSmart® Buffer	10055736	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10055732	Pass	

Assay Name/Specification	Lot # 10059959
Endonuclease Activity (Nicking) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of supercoiled φX174 DNA and a	Pass
minimum of 100 units of Nsil-HF incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of Nsil-HF incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Test (15 minute Digest) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and 1 µl of Nsil-HF incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with Nsil-HF, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments,	Pass



R3127L / Lot: 10059959

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Anthony Francis
Production Scientist

04 Dec 2019

Jay Minichiello

Packaging Quality Control Inspector

27 Dec 2019



R3127L / Lot: 10059959

Page 2 of 2