

## New England Biolabs Certificate of Analysis

**Product Name:** BclI-HF<sup>®</sup>  
**Catalog Number:** R3160S  
**Concentration:** 20,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (dam-) in rCutSmart Buffer in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10168320  
**Expiration Date:** 05/2024  
**Storage Temperature:** -20°C  
**Storage Conditions:** 300 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 500 µg/ml rAlbumin, (pH 7.4 @ 25°C)  
**Specification Version:** PS-R3160S/L v2.0

BclI-HF <sup>®</sup> Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3160SVIAL	BclI-HF <sup>®</sup>	10151996	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10163561	Pass
B6004SVIAL	rCutSmart <sup>™</sup> Buffer	10163560	Pass

Assay Name/Specification	Lot # 10168320
<b>Ligation and Recutting (Terminal Integrity)</b> After a 20-fold over-digestion of Lambda dam- DNA with BclI-HF, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BclI-HF.	Pass
<b>Protein Purity Assay (SDS-PAGE)</b> BclI-HF is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in rCutSmart <sup>™</sup> Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 100 units of BclI-HF incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Functional Testing (15 minute Digest)</b> A 50 µl reaction in rCutSmart <sup>™</sup> Buffer containing 1 µg of Lambda dam- DNA and 1 µl of BclI-HF incubated for 15 minutes at 37°C results in complete digestion as determined	Pass

Assay Name/Specification	Lot # 10168320
<p>by agarose gel electrophoresis.</p> <p><b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda dam- DNA and a minimum of 60 units of BclI-HF incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	<p><b>Pass</b></p>
<p><b>qPCR DNA Contamination (E. coli Genomic)</b> A minimum of 20 units of BclI-HF is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.




---

Penghua Zhang  
Production Scientist  
16 May 2022




---

Erin Varney  
Packaging Quality Control Inspector  
21 Oct 2022