

New England Biolabs Certificate of Analysis

Product Name: Amylose Resin High Flow
Catalog Number: E8022S
Lot Number: 10042731
Expiration Date: 01/2022
Storage Temperature: 4°C
Specification Version: PS-E8022S/L v1.0

Amylose Resin High Flow Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E8022SVIAL	Amylose Resin High Flow	10027735	Pass

Assay Name/Specification	Lot # 10042731
<p>Functional Binding Assay (Resin Binding Capacity) Amylose Resin High Flow (1 ml) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosinΔSal fusion protein (8 ml) was then passed through the column at 25°C, then washed with column buffer and the target protein eluted with 4 ml of column buffer containing 10 mM maltose. Binding capacity was determined to be >4 mg MBP5*-paramyosinΔSal /ml of resin based on A280 of the eluate.</p>	Pass
<p>Functional Binding Assay (Resin Binding Specificity) Amylose Resin High Flow (1 ml) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosinΔSal fusion protein (8 ml) was then passed through the column at 25°C, and then washed with column buffer. The target protein was eluted with 4 ml of column buffer containing 10 mM maltose. SDS-PAGE of the eluate on a 10-20% Tris-Glycine gel confirms low non-specific binding of extract proteins and high isolation of target.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



Brad Landgraf
Production Scientist
08 Jan 2019



Michael Tonello
Packaging Quality Control Inspector
10 Jun 2019