

## New England Biolabs Certificate of Analysis

**Product Name:** SNAP-Surface® Alexa Fluor 546  
**Catalog Number:** S9132S  
**Lot Number:** 10033295  
**Expiration Date:** 01/2024  
**Storage Temperature:** -20°C  
**Specification Version:** PS-S9132S v2.0

SNAP-Surface® Alexa Fluor 546 Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S9132SVIAL	SNAP-Surface® Alexa Fluor® 546	10033296	Pass

Assay Name/Specification	Lot # 10033295
<b>Cellular Protein Labeling (Cell Surface)</b> Mammalian cells transfected with pSNAPf-ADRβ2 expressing Beta-2 adrenergic receptor (cell surface) were labeled with 5 μM SNAP-Surface® Alexa Fluor® 546 for 1 hour and visualized by fluorescence microscopy resulting in the expected cell surface labeling.	<b>Pass</b>
<b>Cellular Protein Labeling (Intracellular)</b> Mammalian cells transfected with pSNAPf-H2B expressing Histone H2B protein (nucleus) were labeled with 5 μM SNAP-Surface® Alexa Fluor® 546 for 1 hour and visualized by fluorescence microscopy resulting in no intracellular labeling.	<b>Pass</b>
<b>Identity (Mass Spectrometry)</b> The observed molecular mass of SNAP-Surface® Alexa Fluor® 546 is 1211.2 Da +/- 1 Da as determined by mass spectrometry analysis.	<b>Pass</b>
<b>In Vitro Protein Labeling</b> A 50 μl reaction in 1X PBS and 1 mM DTT containing 5 μM of SNAP-tag® Purified Protein and a minimum of 10 μM of SNAP-Surface® Alexa Fluor® 546 is incubated for 1 hour at 37°C results in the expected labeled product that is visualized on SDS-PAGE by fluorescent detection.	<b>Pass</b>
<b>Physical Purity (HPLC)</b> SNAP-Surface® Alexa Fluor® 546 is ≥ 90% pure as determined by HPLC analysis.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

*Christopher R. Provost*

---

Chris Provost  
Production Scientist  
29 Jan 2019

*Jay Minichiello*

---

Jay Minichiello  
Packaging Quality Control Inspector  
01 Feb 2019