


New England Biolabs Certificate of Analysis

Product Name: Monarch RNA Purification Columns
 Catalog Number: T2007L
 Packaging Lot Number: 10044092
 Expiration Date: 06/2021
 Storage Temperature: 25°C
 Specification Version: PS-T2007L v1.0

Assay Name/Specification	Lot # 10044092
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® RNA Purification Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
19 Dec 2019



Heidi Church
Packaging Quality Control Inspector
19 Dec 2019