

SCOVERY 240 County Road
ENUINE Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II DNA Library Prep Kit for Illumina®

Catalog Number: E7645L
Packaging Lot Number: 10180056
Expiration Date: 12/2023
Storage Temperature: -20°C

Specification Version: PS-E7645S/L v1.0

NEBNext® Ultra™ II DNA Library Prep Kit for Illumina® Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
E7649AAVIAL	NEBNext® Ultra™ II Q5® Master Mix	10169201	Pass	
E7648AAVIAL	NEBNext® Ultra™ II Ligation Master Mix	10169199	Pass	
E7647AAVIAL	NEBNext® Ultra™ II End Prep Reaction Buffer	10169198	Pass	
E7646AAVIAL	NEBNext® Ultra™ II End Prep Enzyme Mix	10169197	Pass	
E7374AAVIAL	NEBNext® Ligation Enhancer	10169200	Pass	

Assay Name/Specification	Lot # 10180056
Functional Testing (Library Construction, DNA)	Pass
Each set of reagents is functionally validated and compared to a previous lot	
through construction of libraries made from commercially available genomic DNA,	
using the kit's minimum and maximum input requirements. Libraries made from previous	
and current lots for both input DNA amounts are sequenced together on the same	
Illumina® flow cell and compared across various metrics including library yield,	
fraction of reads aligning to the reference, GC bias, and insert size.	
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in NEBNext®	
Ultra™ II DNA Library Prep Kit for Illumina® and meet the designated specifications.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7645L / Lot: 10180056 Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

**Christine Sumner Production Scientist** 09 Feb 2023

Josh Hersey Packaging Quality Control Inspector

09 Feb 2023



E7645L / Lot: 10180056

Page 2 of 2