

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat)
Catalog Number: E7750X
Packaging Lot Number: 10223827
Expiration Date: 01/2026
Storage Temperature: -20°C
Specification Version: PS-E7750S/L/X v1.0

NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7753-4VIAL	NEBNext® DNase I	10223828	Pass
E7752-4VIAL	NEBNext® Thermostable RNase H	10223829	Pass
E7751-4VIAL	NEBNext® Globin & rRNA Depletion Solutio	10223830	Pass
E6317-4VIAL	Nuclease-free Water	10223833	Pass
E6315-4VIAL	DNase I Reaction Buffer	10223832	Pass
E6314-4VIAL	NEBNext® Probe Hybridization Buffer	10223834	Pass
E6312-4VIAL	RNase H Reaction Buffer	10223831	Pass

Assay Name/Specification	Lot # 10223827
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.</p>	Pass
<p>Functional Testing (Globin and rRNA Depletion) The NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christine Sumner

Christine Sumner
Production Scientist
29 Feb 2024

T Monkiewicz

Talia Monkiewicz
Packaging Quality Control Inspector
08 Apr 2024