

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II RNA First Strand Synthesis Module

Catalog Number: E7771S

Lot Number: 10050670

Expiration Date: 09/2020

Storage Temperature: -20°C

Specification Version: PS-E7771S/L v1.0

| NEBNext® Ultra™ II RNA First Strand Synthesis Module Component List |   |            |                      |  |
|---|---|------------|----------------------|--|
| <b>NEB Part Number</b>  | Component Description                           | Lot Number | Individual QC Result |  |
| E7766AVIAL  | NEBNext® Strand Specificity Reagent             | 10039901   | Pass                 |  |
| E7761AVIAL  | NEBNext® First Strand Synthesis Enzyme Mix      | 10039900   | Pass                 |  |
| E7422AVIAL  | Random Primers                                  | 10039899   | Pass                 |  |
| E7421AVIAL  | NEBNext® First Strand Synthesis Reaction Buffer | 10039898   | Pass                 |  |

| Assay Name/Specification   | Lot # 10050670 |
|--|----------------|
| Functional Testing (Library Construction, RNA)  Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from the previous and current lots for both input RNA amounts are sequenced together on the same Illumina | Pass           |
| flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3' transcript coverage, and fraction of reads mapping to a reference.  |                |
| * Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II RNA First Strand Synthesis Module and meet the designated specifications.   | Pass           |

This product has been tested and shown to be in compliance with all specifications.



E7771S / Lot: 10050670 Page 1 of 2 Christin Summ

Christine Sumner Production Scientist 18 Jul 2019 Michael Tonello

Packaging Quality Control Inspector

18 Jul 2019



E7771S / Lot: 10050670

Page 2 of 2