

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Ph.D.™-C7C Phage Display Peptide Library Kit

Catalog Number: E8120S
Packaging Lot Number: 10109412
Expiration Date: 07/2023
Storage Temperature: -20°C

Specification Version: PS-E8120S v2.0

Ph.D.™-C7C Phage Display Peptide Library Kit Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
S1259AVIAL	-96 glll Sequencing Primer (20-mer)	10114711	Pass	
S1258AVIAL	-28 glll Sequencing Primer (22-mer)	10114710	Pass	
N7024AVIAL	Biotin	10111664	Pass	
N7023AVIAL	Streptavidin, lyophilized	10111665	Pass	
E8121AVIAL	Ph.D.™-C7C Phage Display Peptide Library	10109413	Pass	
E4104SVIAL	E.coli K12 ER2738	10055843	Pass	

Assay Name/Specification	Lot # 10109412
Absolute Phage Titer Infection of a mid-log culture of E. coli ER2738 with Ph.D.™-C7C Phage Display	Pass
Peptide Library followed by plating, yields ≥ 1 x 101³ pfu/ml.	
Functional Testing (Panning) A 100-fold representation of the Ph.D.™-C7C Phage Display Peptide Library containing approximately 1011 pfu is diluted in 200 µl TBS and panned against 300 ng of anti-FLAG® monoclonal antibody. The bound phage is affinity captured using magnetic beads and eluted with 1 ml of 0.2M Glycine-HCl, pH 2.2. After three rounds of selection, ≥75% of sequences contain a motif related to the known epitope for the antibody.	Pass
Phage Contamination (Environmental) A 1:100 dilution of an overnight culture of E. coli ER2738 was made in 20 ml LB, to which 10³ pfu of Ph.D.™-C7C Phage Display Peptide Library was added. The flask was incubated at 37°C on a rotating shaker for 5 hours. A 1 ml volume of culture was removed and centrifuged. A volume of culture supernatant equivalent to the initial PFU input was added to a second, 20 ml culture like the first. The final culture supernatant was plated on three LB/IPTG/Xgal plates and then titered. Fewer than 20% clear or white plaques were observed in a minimum of 100 total plaques counted on	Pass



E8120S / Lot: 10109412

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Alicia Bielik

**Production Scientist** 

09 Jul 2021

Josh Hersey

Packaging Quality Control Inspector

09 Jul 2021



E8120S / Lot: 10109412

Page 2 of 2