

*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Supercoiled DNA Ladder
Catalog Number:	N0472S
Concentration:	500 μg/ml
Unit Definition:	N/A
Packaging Lot Number:	10165650
Expiration Date:	10/2024
Storage Temperature:	-20°C
Storage Conditions:	10 mM Tris-HCl (pH 8.0), 1 mM EDTA
Specification Version:	PS-N0472S v1.0

Supercoiled DNA Ladder Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N0472SVIAL	Supercoiled DNA Ladder	10165648	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10158560	Pass	

Assay Name/Specification	Lot # 10165650
<b>A260/A280 Assay</b> The ratio of UV absorption of Supercoiled DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	Pass
<b>Electrophoretic Pattern (Marker)</b> The banding pattern of Supercoiled DNA Ladder on a 0.8% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 $\mu$ g/ml.	Pass
<b>DNA Concentration (A260)</b> The concentration of Supercoiled DNA Ladder is between 500 and 550 $\mu$ g/ml as determined by UV absorption at 260 nm.	Pass
<b>Non-Specific DNase Activity (DNA, 16 hour)</b> A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of Supercoiled DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Mulhiceh

Vanessa Mathieu-Sheltry Production Scientist 29 Sep 2022

Josh Hersey

Packaging Quality Control Inspector 07 Oct 2022

