240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350

www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: BsmI

Catalog #: R0134S/L Concentration: 10,000 units/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 65°C in a total reaction

volume of 50 μ l.

Lot #: 0161504 04/2015 Assay Date: Expiration Date: 04/2017 -20°C Storage Temp:

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 μg/ml BSA

Specification Version: PS-R0134S/L v1.0 Effective Date: 29 Jul 2013

Assay Name/Specification (minimum release criteria)	Lot #0161504
Exonuclease Activity (Radioactivity Release) - A 50 μl reaction in CutSmart TM Buffer containing 1 μg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 100 units of BsmI incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 10-fold over-digestion of Lambda DNA with BsmI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 4 hours at 25°C. Of these ligated fragments, >95% can be recut with BsmI.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 μl reaction in CutSmart TM Buffer containing 1 μg of pBR322 DNA and a minimum of 30 units of BsmI incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Authorized by Derek Robinson 29 Jul 2013

nga ISO 9001





Inspected by Katrina Francescone 25 Mar 2015

^{*} The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (#R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.