

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Smal
Catalog Number: R0141L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA (HindIII digest) in 1 hour at 25°C in a total reaction

volume of 50 μl.

Packaging Lot Number: 10119278
Expiration Date: 09/2023
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 500 μg/ml BSA

Specification Version: PS-R0141S/L v1.0

Smal Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0141LVIAL	Smal	10119277	Pass	
B7024AVIAL	Gel Loading Dye, Purple (6X)	10108733	Pass	
B6004SVIAL	rCutSmart™ Buffer	10115113	Pass	

Assay Name/Specification	Lot # 10119278
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 20 Units of Smal incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda-HindIII DNA and a minimum of 60 Units of Smal incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Smal, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Smal.	Pass
Exonuclease Activity (Radioactivity Release)	Pass



R0141L / Lot: 10119278

Page 1 of 2

Assay Name/Specification	Lot # 10119278
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of Smal incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	
Blue-White Screening (Terminal Integrity) A sample of pUC19 vector linearized with a 10-fold excess of Smal, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Penghua Zhang Production Scientist

14 Sep 2021

Michael Tonello

Packaging Quality Control Inspector

14 Sep 2021



R0141L / Lot: 10119278

Page 2 of 2