

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Acil
Catalog Number: R0551S
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Lot Number: 10044820
Expiration Date: 05/2021
Storage Temperature: -20°C

Storage Conditions: 100 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0551S/L v1.0

| Acil Component List | | | | |
|------------------------|------------------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| R0551SVIAL | Acil | 10044821 | Pass | |
| B7204SVIAL | CutSmart® Buffer | 10042966 | Pass | |
| B7024SVIAL | Gel Loading Dye, Purple (6X) | 10042958 | Pass | |

| Assay Name/Specification | Lot # 10044820 |
|--|----------------|
| Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 30 units of Acil incubated for 4 hours at 37°C releases <0.1% of the total radioactivity. | Pass |
| Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Acil, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~50% can be recut with Acil. | Pass |
| Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 30 Units of Acil incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |

This product has been tested and shown to be in compliance with all specifications.



R0551S / Lot: 10044820 Page 1 of 2 Production Scientist 01 May 2019

Jay Minichiello

Packaging Quality Control Inspector

31 May 2019