

240 County Road Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: BamHI-HF®
Catalog Number: R3136T
Concentration: 100,000 U/mI

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in rCutSmart™ Buffer in 1 hour at 37°C in a total

reaction volume of 50 μl.

Packaging Lot Number: 10168855
Expiration Date: 11/2024
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 50 mM KCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200

μg/ml rAlbumin (pH 7.4 @ 25°C)

Specification Version: PS-R3136T/M v2.0

BamHI-HF® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R3136TVIAL	BamHI-HF®	10168853	Pass	
B7024AVIAL	Gel Loading Dye, Purple (6X)	10163561	Pass	
B6004SVIAL	rCutSmart™ Buffer	10163560	Pass	

Assay Name/Specification	Lot # 10168855
Endonuclease Activity (Nicking) A 50 μl reaction in rCutSmart™ Buffer containing 1 μg of supercoiled PhiX174 DNA and a minimum of 100 units of BamHI-HF® incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BamHI-HF® incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 units of BamHI-HF® incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 20 units of BamHI-HF® is screened for the presence of E. coli genomic	Pass



R3136T / Lot: 10168855

Page 1 of 2

Assay Name/Specification	Lot # 10168855
DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	
Blue-White Screening (Terminal Integrity) A sample of pUC19 vector linearized with a 10-fold excess of BamHI-HF®, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Functional Testing (15 minute Digest) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and 1 µl of BamHI-HF® incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 50-fold over-digestion of Lambda DNA with BamHI-HF®, >95% of the DNA ragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated ragments, >95% can be recut with BamHI-HF®.	Pass
Protein Purity Assay (SDS-PAGE) BamHI-HF® is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

YunJie Sun \
Production Scientist

03 Nov 2022

Josh Hersey

Packaging Quality Control Inspector

07 Nov 2022



R3136T / Lot: 10168855

Page 2 of 2