

be INSPIRED drive DISCOVERY stay GENUINE

320 Newburyport Turnpike Rowley, MA 01969-2002 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Monarch® Plasmid Miniprep Kit
Catalog Number:	T1010S
Packaging Lot Number:	10166553
Expiration Date:	11/2024
Storage Temperature:	25°C
Specification Version:	PS-T1010S/L v2.0

Monarch® Plasmid Miniprep Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T1017-2	Monarch® Plasmid Miniprep Columns	10166535	Pass
T1016-2	Monarch® DNA Elution Buffer	10166531	Pass
T1015-2	Monarch® Plasmid Wash Buffer 2	10166528	Pass
T1014-2	Monarch® Plasmid Wash Buffer 1	10166525	Pass
T1013-2	Monarch® Plasmid Neutralization Buffer (B3)	10166522	Pass
T1012-2	Monarch® Plasmid Lysis Buffer (B2)	10166517	Pass
T1011-2	Monarch® Plasmid Resuspension Buffer (B1)	10166512	Pass

Assay Name/Specification	Lot # 10166553
Functional Testing (Plasmid Purification and Restriction Digestion) pUC19 plasmid is purified from 8 NEB10-Beta samples using the Monarch® Plasmid Miniprep Kit and digested with XmnI resulting in ≥90% conversion to a linearized fragment and no visible gDNA or RNA observed by gel electrophoresis in ≥75% of the samples.	Pass
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch <sup>®</sup> Plasmid Miniprep Columns are tested with 1.5 $\mu$ g of 1 Kb Ladder resulting in $\geq$ 75% recovery in $\geq$ 80% of the samples. OD 260/280 and 260/230 are $\geq$ 1.75 in $\geq$ 80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.





*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

320 Newburyport Turnpike Rowley, MA 01969-2002 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Ana Egana Production Scientist 15 Dec 2022

Tul But

Michael Burzyk Packaging Quality Control Inspector 15 Dec 2022

